

Setup Procedure for Other Flow Cytometers

For setup of flow cytometers not specifically addressed, the setup procedure will differ from one to another. It is very important for the end-user to set up the machine following similar guidelines. **In this case, machine compensation between the reporter and beads classification channels is strongly recommended.**

New flow cytometer setup instructions will be added to our website when they become available.